

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 05/11/2022

OPERATOR NAME:
 LANCE M BARBER
 Permit #: 42841
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:01
INTERNAL STANDARD	VERIFIED	23:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:01
BLANK TEST	0.00	23:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:02
BLANK TEST	0.00	23:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:04
BLANK TEST	0.00	23:05

