

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 05/08/2022

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:56
INTERNAL STANDARD	VERIFIED	00:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:56
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:59
BLANK TEST	0.00	00:59

