

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 05/09/2022

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:24
INTERNAL STANDARD	VERIFIED	07:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:25
BLANK TEST	0.00	07:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:26
BLANK TEST	0.00	07:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:27
BLANK TEST	0.00	07:28

