

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 05/10/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 37635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:51
INTERNAL STANDARD	VERIFIED	12:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:51
BLANK TEST	0.00	12:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:52
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:54
BLANK TEST	0.00	12:54

