

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 05/02/2022

OPERATOR NAME:
JEFFREY E PETERMAN
Permit #: 49748
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:58
INTERNAL STANDARD	VERIFIED	06:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:59
BLANK TEST	0.00	07:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:00
BLANK TEST	0.00	07:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:01
BLANK TEST	0.00	07:02

