

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 05/09/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:27
INTERNAL STANDARD	VERIFIED	22:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:27
BLANK TEST	0.00	22:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:29
BLANK TEST	0.00	22:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:30
BLANK TEST	0.00	22:31

JP
#1203

