

SUPERVISOR TEST

WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 05/17/2022

OPERATOR NAME:
 ERIC C WRIGHT
 Permit #: 49047
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 22030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 20:28 |
| INTERNAL STANDARD | VERIFIED | 20:28 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 20:28 |
| BLANK TEST | 0.00 | 20:29 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 20:30 |
| BLANK TEST | 0.00 | 20:30 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 20:31 |
| BLANK TEST | 0.00 | 20:32 |

