

SUPERVISOR TEST

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 05/20/2022

OPERATOR NAME:  
 LANCE M BARBER  
 Permit #: 42841  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:19
INTERNAL STANDARD	VERIFIED	18:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:19
BLANK TEST	0.00	18:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:21
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:22
BLANK TEST	0.00	18:23

