

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 05/27/2022

OPERATOR NAME:
LANCE M BARBER
Permit #: 42841
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 04:39 |
| INTERNAL STANDARD | VERIFIED | 04:39 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 04:39 |
| BLANK TEST | 0.00 | 04:40 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 04:40 |
| BLANK TEST | 0.00 | 04:41 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 04:42 |
| BLANK TEST | 0.00 | 04:43 |

