

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 06/01/2022

OPERATOR NAME:
DAVID M COSTELLO
Permit #: 40891
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:48
INTERNAL STANDARD	VERIFIED	08:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:48
BLANK TEST	0.00	08:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:50
BLANK TEST	0.00	08:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:51
BLANK TEST	0.00	08:52

