

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 04/30/2022

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:24
INTERNAL STANDARD	VERIFIED	20:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:25
BLANK TEST	0.00	20:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:26
BLANK TEST	0.00	20:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:27
BLANK TEST	0.00	20:28

