

SOLUTION CHANGE

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 05/01/2022

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:51
INTERNAL STANDARD	VERIFIED	06:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:54
BLANK TEST	0.00	06:55

