

SOLUTION CHANGE

-----  
 GRANVILLE VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 113306  
 -----

Date: 05/30/2022

OPERATOR NAME:  
 JASON VANDENBURGH  
 Permit #: 37635  
 Operator Agency/Dept: GRANVILLE VILLAGE PD  
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:03
INTERNAL STANDARD	VERIFIED	08:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:03
BLANK TEST	0.00	08:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:04
BLANK TEST	0.00	08:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:06
BLANK TEST	0.00	08:06

