

SUPERVISOR TEST

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 06/06/2022

OPERATOR NAME:  
JASON M VANDENBURGH  
Permit #: 37635  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:46
INTERNAL STANDARD	VERIFIED	06:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:46
BLANK TEST	0.00	06:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:47
BLANK TEST	0.00	06:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:48
BLANK TEST	0.00	06:49

