

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 06/06/2022

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:02
INTERNAL STANDARD	VERIFIED	17:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:02
BLANK TEST	0.00	17:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:03
BLANK TEST	0.00	17:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:04
BLANK TEST	0.00	17:05

