

SUPERVISOR TEST

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 06/13/2022

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:45
INTERNAL STANDARD	VERIFIED	08:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:45
BLANK TEST	0.00	08:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:47
BLANK TEST	0.00	08:48

