

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 06/06/2022

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:35
INTERNAL STANDARD	VERIFIED	17:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:36
BLANK TEST	0.00	17:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:37
BLANK TEST	0.00	17:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:38
BLANK TEST	0.00	17:39

Handwritten signature and initials: #691

