

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 06/13/2022

OPERATOR NAME:
TREVOR J PECK
Permit #: 49747
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:41
INTERNAL STANDARD	VERIFIED	20:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:42
BLANK TEST	0.00	20:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:43
BLANK TEST	0.00	20:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:44
BLANK TEST	0.00	20:45

6/13
#637

