

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 06/08/2022

OPERATOR NAME:
 LANCE M BARBER
 Permit #: 42841
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:13
INTERNAL STANDARD	VERIFIED	20:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:13
BLANK TEST	0.00	20:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:15
BLANK TEST	0.00	20:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:16
BLANK TEST	0.00	20:17

