

SUPERVISOR TEST

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 06/22/2022

OPERATOR NAME:
 EVAN J HOAGLAND
 Permit #: 49911
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:25
INTERNAL STANDARD	VERIFIED	10:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:25
BLANK TEST	0.00	10:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:27
BLANK TEST	0.00	10:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:28
BLANK TEST	0.00	10:29

