

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 07/07/2022

OPERATOR NAME:  
LANCE M BARBER  
Permit #: 42841  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:44
INTERNAL STANDARD	VERIFIED	19:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:44
BLANK TEST	0.00	19:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:45
BLANK TEST	0.00	19:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:47
BLANK TEST	0.00	19:48

