

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 06/18/2022

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:54
INTERNAL STANDARD	VERIFIED	19:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:55
BLANK TEST	0.00	19:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:56
BLANK TEST	0.00	19:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:57
BLANK TEST	0.00	19:58

