

SUPERVISOR TEST

 WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 06/25/2022

OPERATOR NAME:
 ALEXANDER T GILMORE
 Permit #: 47649
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:02
INTERNAL STANDARD	VERIFIED	18:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:03
BLANK TEST	0.00	18:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:04
BLANK TEST	0.00	18:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:05
BLANK TEST	0.00	18:06

