

SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 06/06/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:09
INTERNAL STANDARD	VERIFIED	23:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:10
BLANK TEST	0.00	23:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:11
BLANK TEST	0.00	23:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:12
BLANK TEST	0.00	23:13

JD
#1203

