

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 06/13/2022

OPERATOR NAME:  
JOSEPH P DEANGELO  
Permit #: 49218  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:02
INTERNAL STANDARD	VERIFIED	17:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:03
BLANK TEST	0.00	17:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:04
BLANK TEST	0.00	17:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:05
BLANK TEST	0.00	17:06

*JD*  
#1203

