

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 06/27/2022

OPERATOR NAME:
JEFFREY E PETERMAN
Permit #: 49748
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:53
INTERNAL STANDARD	VERIFIED	04:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:54
BLANK TEST	0.00	04:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:55
BLANK TEST	0.00	04:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:56
BLANK TEST	0.00	04:57

