

SOLUTION CHANGE

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 07/01/2022

OPERATOR NAME:  
PAUL S HENRY  
Permit #: 49918  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:56
INTERNAL STANDARD	VERIFIED	23:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:56
BLANK TEST	0.00	23:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:58
BLANK TEST	0.00	23:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:59
BLANK TEST	0.00	00:00

