

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 06/27/2022

OPERATOR NAME:
 ALIX A MESSINA
 Permit #: 50329
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:38
INTERNAL STANDARD	VERIFIED	07:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:39
BLANK TEST	0.00	07:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:40
BLANK TEST	0.00	07:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:41
BLANK TEST	0.00	07:42

