

SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/05/2022

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:35
INTERNAL STANDARD	VERIFIED	07:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:35
BLANK TEST	0.00	07:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:37
BLANK TEST	0.00	07:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:38
BLANK TEST	0.00	07:39

