

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 07/19/2022

OPERATOR NAME:  
JOSEPH P DEANGELO  
Permit #: 49218  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:32
INTERNAL STANDARD	VERIFIED	15:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:33
BLANK TEST	0.00	15:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:34
BLANK TEST	0.00	15:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:35
BLANK TEST	0.00	15:36

*Handwritten signature and number:*  
#1203

