

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 07/25/2022

OPERATOR NAME:  
JOSEPH P DEANGELO  
Permit #: 49218  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:48
INTERNAL STANDARD	VERIFIED	17:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:48
BLANK TEST	0.00	17:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:50
BLANK TEST	0.00	17:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:51
BLANK TEST	0.00	17:52

JD  
#1203

