

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 07/24/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 37635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:19
INTERNAL STANDARD	VERIFIED	08:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:20
BLANK TEST	0.00	08:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:21
BLANK TEST	0.00	08:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:22
BLANK TEST	0.00	08:23

