

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 07/11/2022

OPERATOR NAME:
TREVOR J PECK
Permit #: 49747
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:27
INTERNAL STANDARD	VERIFIED	02:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:27
BLANK TEST	0.00	02:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:28
BLANK TEST	0.00	02:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:29
BLANK TEST	0.00	02:30

#637

7/11/22

