

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 08/29/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:18
INTERNAL STANDARD	VERIFIED	22:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:19
BLANK TEST	0.00	22:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:20
BLANK TEST	0.00	22:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:21
BLANK TEST	0.00	22:22

JD
#1203

