

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 08/15/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:21
INTERNAL STANDARD	VERIFIED	20:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:21
BLANK TEST	0.00	20:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:23
BLANK TEST	0.00	20:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:24
BLANK TEST	0.00	20:25

JJD
#1203

