

SOLUTION CHANGE

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 08/01/2022

OPERATOR NAME:  
JOSEPH P DEANGELO  
Permit #: 49218  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:00
INTERNAL STANDARD	VERIFIED	22:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:01
BLANK TEST	0.00	22:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04

JD  
#1203

