

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 08/22/2022

OPERATOR NAME:
 LANCE M BARBER
 Permit #: 42841
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:03
INTERNAL STANDARD	VERIFIED	20:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:04
BLANK TEST	0.00	20:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:05
BLANK TEST	0.00	20:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:06
BLANK TEST	0.00	20:07

