

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 08/20/2022

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 22190

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 23:17 |
| INTERNAL STANDARD | VERIFIED | 23:18 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 23:18 |
| BLANK TEST | 0.00 | 23:19 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 23:19 |
| BLANK TEST | 0.00 | 23:20 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 23:20 |
| BLANK TEST | 0.00 | 23:21 |

