

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 08/28/2022

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:45
INTERNAL STANDARD	VERIFIED	15:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:45
BLANK TEST	0.00	15:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:47
BLANK TEST	0.00	15:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:48
BLANK TEST	0.00	15:49

