

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 08/22/2022

OPERATOR NAME:
ERNEST R BASSETT
Permit #: 25518
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:24
INTERNAL STANDARD	VERIFIED	06:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:25
BLANK TEST	0.00	06:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:26
BLANK TEST	0.00	06:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:27
BLANK TEST	0.00	06:28

