

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 08/15/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 37635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:23
INTERNAL STANDARD	VERIFIED	07:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:24
BLANK TEST	0.00	07:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:25
BLANK TEST	0.00	07:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:26
BLANK TEST	0.00	07:27

