

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 08/15/2022

OPERATOR NAME:  
LANCE M BARBER  
Permit #: 42841  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:51
INTERNAL STANDARD	VERIFIED	05:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:51
BLANK TEST	0.00	05:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:52
BLANK TEST	0.00	05:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:54
BLANK TEST	0.00	05:55

