

SUPERVISOR TEST

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 07/21/2022

OPERATOR NAME:  
 LANCE M BARBER  
 Permit #: 42841  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:38
INTERNAL STANDARD	VERIFIED	23:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:39
BLANK TEST	0.00	23:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:42

