

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 08/15/2022

OPERATOR NAME:
DAVID F BUXTON
Permit #: 33398
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:57
INTERNAL STANDARD	VERIFIED	07:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:57
BLANK TEST	0.00	07:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:59
BLANK TEST	0.00	08:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:00
BLANK TEST	0.00	08:01

