

SUPERVISOR TEST

WHITEHALL VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 113806

Date: 07/09/2022

OPERATOR NAME:  
 BRYAN J GRECO  
 Permit #: 44077  
 Operator Agency/Dept: WHITEHALL VILL. P.D.  
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:10
INTERNAL STANDARD	VERIFIED	18:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:11
BLANK TEST	0.00	18:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:12
BLANK TEST	0.00	18:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:13
BLANK TEST	0.00	18:14

