

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/25/2022

OPERATOR NAME:
WILLIAM M GREENE
Permit #: 48760
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:59
INTERNAL STANDARD	VERIFIED	23:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:59
BLANK TEST	0.00	00:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:00
BLANK TEST	0.00	00:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:01
BLANK TEST	0.00	00:02

