

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/28/2022

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:37
INTERNAL STANDARD	VERIFIED	19:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:37
BLANK TEST	0.00	19:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:38
BLANK TEST	0.00	19:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:40
BLANK TEST	0.00	19:41

[Handwritten Signature] #69

