

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 08/08/2022

OPERATOR NAME:  
ROBERT J SULLIVAN  
Permit #: 34106  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:39
INTERNAL STANDARD	VERIFIED	09:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:40
BLANK TEST	0.00	09:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:41
BLANK TEST	0.00	09:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:42
BLANK TEST	0.00	09:43

