

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/21/2022

OPERATOR NAME:
JASON R DIAMOND
Permit #: 37496
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:08
INTERNAL STANDARD	VERIFIED	01:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:08
BLANK TEST	0.00	01:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:11
BLANK TEST	0.00	01:12

