

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 08/29/2022

OPERATOR NAME:  
JASON R DIAMOND  
Permit #: 37496  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:44
INTERNAL STANDARD	VERIFIED	20:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:44
BLANK TEST	0.00	20:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:46
BLANK TEST	0.00	20:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:47
BLANK TEST	0.00	20:48

